

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000187  
13:12 17/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:41  
STERILIZING 3:39  
136.0 C 222kPa 3:39  
136.3 C 222kPa 4:09  
136.3 C 224kPa 4:39  
136.4 C 225kPa 5:09  
136.4 C 225kPa 5:39  
136.3 C 225kPa 6:09  
136.4 C 222kPa 6:39  
136.2 C 224kPa 7:09  
136.2 C 224kPa 7:09  
VENTING 7:09  
AIR DRYING 8:00  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
8F1F06AD09A2C492  
-----