

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003819  
11:12 15/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:41  
STERILIZING 3:41  
136.0 C 220kPa 3:41  
136.2 C 223kPa 4:11  
136.1 C 222kPa 4:41  
136.3 C 224kPa 5:11  
136.1 C 222kPa 5:41  
136.4 C 224kPa 6:11  
136.0 C 219kPa 6:41  
136.4 C 221kPa 7:11  
136.4 C 221kPa 7:11  
VENTING 7:11  
AIR DRYING 8:08  
CYCLE COMPLETE 10:00

-----  
Digital Signature #  
FE74014AF73F93C1  
-----