
Midmark M11 - v1.0.5
Total Cycles: 637

Name: RITTER2
Sterilizer ID: V2313228
12/15/2021 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 184.7 F 3.9
4:00 192.8 F 8.9
6:00 213.4 F 7.6
8:00 227.6 F 9.2
10:00 234.1 F 9.5
12:00 243.2 F 13.2
14:00 243.8 F 12.6
16:00 257.1 F 19.7
18:00 268.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.2 F 29.8
2:00 272.2 F 29.3
3:00 271.5 F 29.3
4:00 272.6 F 29.8

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 222.9 F 0.0
10:00 229.3 F 0.0
15:00 219.1 F 0.0
20:00 216.8 F 0.0
25:00 219.4 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

12/15/2021 02:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____