
Midmark M11 - v1.0.5
Total Cycles: 4046

Name: RITTER2
Sterilizer ID: V2313228
10/22/2024 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.2 F 0.0
2:00 110.1 F 0.0
4:00 119.3 F 0.0
6:00 155.1 F 0.2
8:00 180.4 F 3.0
10:00 194.9 F 6.7
12:00 212.5 F 10.5
14:00 224.0 F 7.9
16:00 229.9 F 9.5
18:00 237.8 F 10.7
20:00 241.2 F 11.8
22:00 239.0 F 10.2
24:00 251.5 F 16.3
26:00 262.3 F 22.6
28:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.1 F 29.4
2:00 272.2 F 28.9
3:00 271.7 F 29.1
4:00 272.5 F 29.3

Min 271.4 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 209.0 F 0.0
10:00 202.1 F 0.0
15:00 194.4 F 0.0
20:00 209.3 F 0.0
25:00 226.4 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:22
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

10/22/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____