

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4900

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2024 12:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.1 F 0.0  
2:00 172.8 F 0.3  
4:00 190.3 F 2.9  
6:00 213.8 F 7.7  
8:00 225.5 F 13.0  
10:00 230.4 F 8.5  
12:00 240.2 F 10.9  
14:00 242.6 F 12.3  
16:00 241.8 F 11.2  
18:00 255.5 F 18.1  
20:00 267.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 29.0  
2:00 272.1 F 28.4  
3:00 271.3 F 28.3  
4:00 272.5 F 29.0  
  
Min 271.1 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 239.2 F 0.0  
10:00 218.8 F 0.0  
15:00 215.2 F 0.0  
20:00 218.6 F 0.0  
25:00 243.1 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:29  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 00:59:58  
CYCLE COMPLETE

09/12/2024 01:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_