
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002376
11:33 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:09
STERILIZING 3:12
136.0 C 220kPa 3:12
136.4 C 225kPa 3:42
136.3 C 225kPa 4:12
136.4 C 226kPa 4:42
136.4 C 226kPa 5:12
136.2 C 224kPa 5:42
136.4 C 225kPa 6:12
136.4 C 226kPa 6:42
136.4 C 226kPa 6:42
VENTING 6:42
AIR DRYING 7:33
CYCLE COMPLETE 25:12

Digital Signature #
C4ACCCD746788567
