

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001992  
10:02 09/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:46  
PRESSURIZING 1:51  
STERILIZING 2:59  
136.4 C 226kPa 2:59  
136.4 C 226kPa 3:29  
136.4 C 226kPa 3:59  
136.4 C 226kPa 4:29  
136.3 C 225kPa 4:59  
136.4 C 226kPa 5:29  
136.3 C 225kPa 5:59  
136.3 C 225kPa 6:29  
136.3 C 225kPa 6:29  
VENTING 6:29  
AIR DRYING 7:21  
CYCLE COMPLETE 7:58  
-----

Digital Signature #  
F9B5939D8B941D7B  
-----