
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001456
10:21 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:48
136.4 C 224kPa 3:48
136.2 C 223kPa 4:18
136.4 C 224kPa 4:48
136.4 C 225kPa 5:18
136.2 C 224kPa 5:48
136.1 C 222kPa 6:18
136.4 C 225kPa 6:48
136.4 C 225kPa 7:18
136.4 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 23:08

Digital Signature #
3E6CA951DB9E41D8
