
Midmark M11 - v1.0.5
Total Cycles: 1136

Name: RITTER2
Sterilizer ID: V2313228
05/30/2022 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.5 F 0.0
2:00 193.0 F 3.3
4:00 195.5 F 8.1
6:00 215.4 F 13.5
8:00 227.6 F 8.6
10:00 236.1 F 11.2
12:00 241.1 F 12.1
14:00 240.5 F 11.1
16:00 254.3 F 18.0
18:00 265.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.5
2:00 271.5 F 28.8
3:00 272.7 F 29.7
4:00 272.1 F 29.1

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 216.6 F 0.0
10:00 227.5 F 0.0
15:00 225.9 F 0.0
20:00 220.0 F 0.0
25:00 217.0 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

05/30/2022 11:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____