

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3849

Name: RITTER 1  
Sterilizer ID: V2178198  
10/02/2023 12:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
2:00 189.4 F 3.7  
4:00 205.7 F 8.9  
6:00 218.3 F 7.8  
8:00 230.7 F 9.2  
10:00 237.9 F 9.9  
12:00 243.7 F 12.4  
14:00 242.6 F 11.2  
16:00 256.0 F 18.1  
18:00 268.0 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 271.8 F 27.9  
2:00 272.4 F 28.3  
3:00 272.6 F 28.7  
4:00 271.6 F 28.0  
  
Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 216.8 F 0.0  
10:00 230.7 F 0.0  
15:00 229.0 F 0.0  
20:00 223.5 F 0.0  
25:00 219.9 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

10/02/2023 01:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_