
Midmark M11 - v1.0.5
Total Cycles: 3019

Name: RITTER2
Sterilizer ID: V2313228
10/17/2023 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.5 F 0.0
2:00 160.6 F 3.2
4:00 182.1 F 7.3
6:00 200.9 F 12.5
8:00 216.3 F 7.0
10:00 237.0 F 12.6
12:00 237.3 F 10.0
14:00 236.2 F 8.6
16:00 250.4 F 15.9
18:00 262.8 F 23.1
20:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.1 F 29.4
2:00 272.8 F 29.8
3:00 271.7 F 29.6
4:00 272.2 F 29.4

Min 271.3 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 254.0 F 0.0
10:00 211.0 F 0.0
15:00 206.0 F 0.0
20:00 200.2 F 0.0
25:00 203.2 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

10/17/2023 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____