

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3006

Name: RITTER2  
Sterilizer ID: V2313228  
10/13/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
2:00 176.4 F 2.1  
4:00 185.5 F 6.6  
6:00 208.0 F 11.8  
8:00 221.9 F 7.2  
10:00 235.5 F 13.0  
12:00 237.9 F 11.0  
14:00 237.8 F 10.0  
16:00 252.4 F 17.1  
18:00 264.6 F 24.2  
20:00 270.9 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.4  
1:00 272.7 F 30.1  
2:00 272.4 F 29.5  
3:00 272.2 F 29.9  
4:00 272.8 F 29.8  
  
Min 271.7 F 29.2  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 220.7 F 0.0  
10:00 206.6 F 0.0  
15:00 203.8 F 0.0  
20:00 201.7 F 0.0  
25:00 199.5 F 0.0  
30:00 198.1 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 20:37  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:41  
CYCLE COMPLETE

10/13/2023 10:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_