

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4253

Name: RITTER2  
Sterilizer ID: V2313228  
01/09/2025 05:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.2 F 0.0  
2:00 119.1 F 0.0  
4:00 140.2 F 0.2  
6:00 178.6 F 0.7  
8:00 203.1 F 4.4  
10:00 216.7 F 8.8  
12:00 229.3 F 13.0  
14:00 233.5 F 9.4  
16:00 238.5 F 9.9  
18:00 244.7 F 13.2  
20:00 243.7 F 12.3  
22:00 256.8 F 19.2  
24:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 271.8 F 28.5  
2:00 272.0 F 28.6  
3:00 272.4 F 28.9  
4:00 272.7 F 29.2

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.6 F 0.0  
5:00 200.6 F 0.0  
10:00 193.0 F 0.0  
15:00 190.6 F 0.0  
20:00 188.7 F 0.0  
25:00 188.5 F 0.0  
30:00 186.9 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:08  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:03:16  
CYCLE COMPLETE

01/09/2025 06:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_