

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4457

Name: RITTER 1  
Sterilizer ID: V2178198  
04/15/2024 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.0 F 0.0  
2:00 157.1 F 2.3  
4:00 167.7 F 5.6  
6:00 185.5 F 9.2  
8:00 212.1 F 9.0  
10:00 219.7 F 6.8  
12:00 230.7 F 11.3  
14:00 232.2 F 7.4  
16:00 244.8 F 13.0  
18:00 242.0 F 11.2  
20:00 253.8 F 17.2  
22:00 264.4 F 23.4  
24:00 271.2 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 271.7 F 28.6  
2:00 272.5 F 28.8  
3:00 272.5 F 29.1  
4:00 271.5 F 29.0

Min 271.2 F 28.5  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.4 F 0.0  
5:00 198.8 F 0.0  
10:00 192.0 F 0.0  
15:00 193.0 F 0.0  
20:00 190.0 F 0.0  
25:00 189.0 F 0.0  
30:00 188.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 24:42  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:03:28  
CYCLE COMPLETE

04/15/2024 12:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_