
Midmark M11 - v1.0.5
Total Cycles: 2202

Name: RITTER2
Sterilizer ID: V2313228
03/31/2023 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.3 F 0.0
2:00 80.7 F 0.0
4:00 88.5 F 0.0
6:00 122.8 F 0.0
8:00 156.8 F 2.8
10:00 201.5 F 6.8
12:00 217.3 F 11.7
14:00 219.3 F 6.4
16:00 231.1 F 10.8
18:00 235.4 F 7.7
20:00 244.2 F 13.2
22:00 243.6 F 12.0
24:00 256.1 F 18.6
26:00 267.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.5 F 28.5
2:00 271.5 F 28.6
3:00 271.6 F 28.6
4:00 271.5 F 28.5

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 234.5 F 0.0
Dry Cycle Aborted

FILLING: 1:05
HEATING: 27:18
STERILIZING: 4:00
VENTING: 2:44
DRYING: 0:12
TOTAL CYCLE: 00:35:19

03/31/2023 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____