

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4232

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2024 10:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.0 F	0.0
2:00	71.4 F	0.0
4:00	76.9 F	0.0
6:00	89.8 F	0.0
8:00	128.7 F	0.1
10:00	156.0 F	0.5
12:00	183.1 F	3.2
14:00	206.0 F	6.1
16:00	215.0 F	9.2
18:00	224.9 F	12.6
20:00	224.6 F	7.4
22:00	235.6 F	11.6
24:00	232.9 F	8.5
26:00	242.7 F	12.3
28:00	243.4 F	12.4
30:00	254.3 F	18.1
32:00	264.0 F	24.1
34:00	272.7 F	29.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.8 F	29.3
1:00	272.9 F	29.8
2:00	272.0 F	29.4
3:00	272.9 F	29.8
4:00	271.8 F	29.3

Min 271.8 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	207.8 F	0.0
	5:00	208.3 F	0.0
	10:00	213.4 F	0.0
	15:00	211.9 F	0.0
	20:00	223.7 F	0.0
Drying Complete	25:00	229.0 F	0.0
	30:00	228.4 F	0.0

FILLING: 0:49  
HEATING: 34:15  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:12:47  
CYCLE COMPLETE

12/20/2024 11:31 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_