
Midmark M11 - v1.0.5
Total Cycles: 2588

Name: RITTER2
Sterilizer ID: V2313228
06/29/2023 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.5 F 0.0
2:00 191.7 F 4.2
4:00 203.7 F 9.2
6:00 221.2 F 7.9
8:00 234.2 F 9.6
10:00 236.6 F 9.4
12:00 245.4 F 13.5
14:00 244.9 F 12.9
16:00 257.6 F 19.7
18:00 268.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.3 F 29.5
2:00 272.2 F 28.9
3:00 271.6 F 29.1
4:00 272.3 F 29.4

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 235.5 F 0.0
10:00 215.6 F 0.0
15:00 194.9 F 0.0
20:00 184.9 F 0.0
25:00 180.7 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

06/29/2023 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____