
Midmark M11 - v1.0.5
Total Cycles: 4245

Name: RITTER2
Sterilizer ID: V2313228
01/07/2025 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	160.5 F	0.0
2:00	158.1 F	0.4
4:00	163.0 F	2.8
6:00	187.3 F	7.6
8:00	210.5 F	10.1
10:00	230.9 F	10.0
12:00	233.2 F	8.1
14:00	239.7 F	10.1
16:00	249.1 F	15.0
18:00	262.2 F	22.5

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 249F
CHAMBER PRESS 14.6PSI
** ITEMS NOT STERILE **

01/07/2025 03:00 PM