
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000151
13:57 14/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:01
PRESSURIZING 4:02
STERILIZING 5:31
136.0 C 222kPa 5:31
136.4 C 226kPa 6:01
136.4 C 225kPa 6:31
136.4 C 223kPa 7:01
136.4 C 226kPa 7:31
136.4 C 225kPa 8:01
136.1 C 223kPa 8:31
136.4 C 226kPa 9:01
136.4 C 226kPa 9:01
VENTING 9:01
AIR DRYING 10:05
CYCLE COMPLETE 10:16

Digital Signature #
998FBBE98B7E8B3C
