
Midmark M11 - v1.0.0
Total Cycles: 4385

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2024 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.9 F 0.0
2:00 110.7 F 0.7
4:00 119.8 F 2.7
6:00 145.1 F 5.6
8:00 169.6 F 9.4
10:00 203.3 F 9.9
12:00 211.3 F 6.6
14:00 222.5 F 10.7
16:00 228.5 F 7.1
18:00 238.1 F 11.3
20:00 235.2 F 8.8
22:00 248.3 F 14.9
24:00 259.2 F 21.0
26:00 269.6 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.7 F 29.5
2:00 272.6 F 29.7
3:00 271.5 F 29.1
4:00 272.3 F 29.3

Min 271.5 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 228.9 F 0.0
10:00 238.7 F 0.0
15:00 230.7 F 0.0
20:00 225.5 F 0.0
25:00 227.5 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:34
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:06:05
CYCLE COMPLETE

03/20/2024 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____