
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004252
10:42 29/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:06
STERILIZING 4:01
136.0 C 220kPa 4:01
136.4 C 223kPa 4:31
136.4 C 226kPa 5:01
136.3 C 225kPa 5:31
136.4 C 226kPa 6:01
136.4 C 223kPa 6:31
136.3 C 224kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 9:27
CYCLE COMPLETE 9:42

Digital Signature #
5DA0C89775504D66
