
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005364
8:32 18/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:53
STERILIZING 2:45
136.0 C 220kPa 2:45
136.4 C 224kPa 3:15
136.0 C 222kPa 3:45
136.4 C 224kPa 4:15
136.4 C 223kPa 4:45
136.4 C 224kPa 5:15
136.2 C 223kPa 5:45
136.4 C 223kPa 6:15
136.4 C 223kPa 6:15
VENTING 6:15
AIR DRYING 7:05
CYCLE COMPLETE 7:09

Digital Signature #
BB7C0E140DC6DB45
