
Midmark M11 - v1.0.0
Total Cycles: 2961

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 180.4 F 3.3
4:00 187.2 F 7.4
6:00 207.2 F 12.3
8:00 220.3 F 6.9
10:00 233.2 F 12.2
12:00 234.4 F 9.2
14:00 238.2 F 9.5
16:00 247.1 F 13.9
18:00 259.4 F 20.6
20:00 270.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.8 F 28.5
3:00 272.0 F 29.0
4:00 272.5 F 29.1

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 224.9 F 0.0
10:00 224.6 F 0.0
15:00 224.7 F 0.0
20:00 225.7 F 0.0
25:00 224.8 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

03/01/2023 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____