

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2578

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2023 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.0 F 0.0  
2:00 163.8 F 0.0  
4:00 169.2 F 2.1  
6:00 199.7 F 6.8  
8:00 215.8 F 12.3  
10:00 226.4 F 7.6  
12:00 238.6 F 13.0  
14:00 238.2 F 10.8  
16:00 237.0 F 9.5  
18:00 251.0 F 16.4  
20:00 263.2 F 23.4  
22:00 271.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.4 F 29.4  
2:00 271.6 F 29.2  
3:00 272.4 F 29.4  
4:00 271.4 F 29.0

Min 271.4 F 28.9  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 201.6 F 0.0  
10:00 208.9 F 0.0  
15:00 213.2 F 0.0  
20:00 216.4 F 0.0  
25:00 212.3 F 0.0  
30:00 210.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:07  
CYCLE COMPLETE

06/28/2023 09:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_