
Midmark M11 - v1.0.0
Total Cycles: 742

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2021 12:09 PM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 166.0 F 2.5
4:00 169.0 F 6.2
6:00 208.9 F 2.0
8:00 215.5 F 6.9
10:00 225.4 F 4.8
12:00 226.2 F 4.4
14:00 238.6 F 9.5
16:00 250.4 F 15.3

STERILIZING
mm:ss Degrees PSI
0:00 251.4 F 15.8
1:00 252.4 F 16.1
2:00 252.6 F 16.5
3:00 251.7 F 15.9
4:00 252.7 F 16.4
5:00 251.9 F 16.0
6:00 251.7 F 16.2
7:00 252.5 F 16.3
8:00 251.8 F 15.9
9:00 251.6 F 16.2
10:00 252.8 F 16.4
11:00 252.1 F 16.1
12:00 251.5 F 15.8
13:00 251.4 F 16.1
14:00 252.9 F 16.5
15:00 252.5 F 16.3
16:00 252.1 F 16.1
17:00 251.6 F 15.9
18:00 251.3 F 15.8
19:00 251.7 F 16.4
20:00 253.0 F 16.5
21:00 252.8 F 16.4
22:00 252.0 F 16.2
23:00 251.8 F 16.1
24:00 251.3 F 15.9
25:00 251.3 F 15.8
26:00 251.4 F 16.0
27:00 252.1 F 16.5
28:00 252.9 F 16.5
29:00 252.8 F 16.4
30:00 252.7 F 16.3

Min 251.0 F 15.7
Max 253.1 F 16.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 224.0 F 0.0
10:00 227.4 F 0.0
15:00 231.2 F 0.0
20:00 183.5 F 0.0
25:00 180.7 F 0.0
Dry Cycle Aborted

FILLING: 1:04
HEATING: 16:45
STERILIZING: 30:00

VENTING: 2:49
DRYING: 29:05
TOTAL CYCLE: 01:19:43

03/24/2021 01:28 PM
ACCEPTED: X
REJECTED:
Operator: U01