

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2580

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2023 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
2:00 179.1 F 4.0  
4:00 185.6 F 8.3  
6:00 208.7 F 12.5  
8:00 223.1 F 8.5  
10:00 237.6 F 11.6  
12:00 238.5 F 11.4  
14:00 238.0 F 10.1  
16:00 251.9 F 16.9  
18:00 263.6 F 23.9  
20:00 271.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.6 F 29.9  
2:00 271.7 F 29.2  
3:00 271.8 F 29.7  
4:00 272.5 F 29.8

Min 271.2 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 213.3 F 0.0  
10:00 234.8 F 0.0  
15:00 224.8 F 0.0  
20:00 226.8 F 0.0  
25:00 226.9 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

06/28/2023 12:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_