
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001210
15:03 18/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:30
136.4 C 225kPa 3:30
136.4 C 225kPa 4:00
136.4 C 221kPa 4:30
136.4 C 225kPa 5:00
136.3 C 224kPa 5:30
136.3 C 223kPa 6:00
136.4 C 223kPa 6:30
136.3 C 224kPa 7:00
136.3 C 224kPa 7:00
VENTING 7:00
AIR DRYING 7:52
CYCLE COMPLETE 7:54

Digital Signature #
B812C4E4160B1C2E
