

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 001592  
11:52 11/18/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 4:37  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 4:37  
-----

Digital Signature #  
F913E96FD4E5E6A9  
-----