
Midmark M11 - v1.0.5
Total Cycles: 2001

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 158.2 F 1.3
4:00 181.2 F 5.2
6:00 200.0 F 10.2
8:00 217.3 F 5.8
10:00 227.6 F 10.6
12:00 230.3 F 7.4
14:00 242.0 F 12.4
16:00 244.0 F 13.0
18:00 256.7 F 19.8
20:00 268.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.4 F 29.9
2:00 271.7 F 29.6
3:00 272.5 F 29.8
4:00 271.5 F 29.4

Min 271.5 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 210.5 F 0.0
10:00 219.7 F 0.0
15:00 226.6 F 0.0
20:00 229.1 F 0.0
25:00 225.0 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

02/02/2023 03:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____