

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2502

Name: RITTER 1  
Sterilizer ID: V2178198  
11/02/2022 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.0 F 0.0  
2:00 148.0 F 2.8  
4:00 162.3 F 6.4  
6:00 185.1 F 10.7  
8:00 215.3 F 4.9  
10:00 222.6 F 9.7  
12:00 232.4 F 8.6  
14:00 237.1 F 11.3  
16:00 236.4 F 8.9  
18:00 248.5 F 15.1  
20:00 259.6 F 21.4  
22:00 270.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.6 F 29.6  
2:00 272.0 F 29.7  
3:00 272.4 F 29.5  
4:00 272.1 F 29.8

Min 271.5 F 29.2  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 237.5 F 0.0  
10:00 227.6 F 0.0  
15:00 226.4 F 0.0  
20:00 227.7 F 0.0  
25:00 231.7 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

11/02/2022 02:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_