

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4418

Name: RITTER 1  
Sterilizer ID: V2178198  
03/28/2024 03:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 106.6 F 0.0  
2:00 106.1 F 0.9  
4:00 133.8 F 3.3  
6:00 155.2 F 6.5  
8:00 215.3 F 10.7  
10:00 215.0 F 3.2  
12:00 222.3 F 6.8  
14:00 232.7 F 10.9  
16:00 228.9 F 6.6  
18:00 242.1 F 12.0  
20:00 238.6 F 9.6  
22:00 251.1 F 15.7  
24:00 262.2 F 22.0  
26:00 272.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 28.8  
2:00 272.3 F 29.1  
3:00 271.9 F 28.6  
4:00 272.8 F 29.2

Min 271.4 F 28.4  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 233.3 F 0.0  
10:00 231.0 F 0.0  
15:00 225.8 F 0.0  
20:00 219.2 F 0.0  
25:00 223.0 F 0.0  
30:00 221.0 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 27:16  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:05:47  
CYCLE COMPLETE

03/28/2024 04:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_