

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2026

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2022 12:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 182.5 F 3.5  
4:00 191.8 F 7.8  
6:00 217.1 F 13.4  
8:00 227.1 F 8.2  
10:00 240.6 F 12.7  
12:00 239.7 F 11.3  
14:00 238.4 F 9.7  
16:00 252.8 F 16.8  
18:00 264.5 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.2 F 28.8  
2:00 272.0 F 28.9  
3:00 271.8 F 28.5  
4:00 272.5 F 29.2  
  
Min 271.4 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 218.9 F 0.0  
10:00 228.5 F 0.0  
15:00 225.5 F 0.0  
20:00 230.0 F 0.0  
25:00 239.8 F 0.0  
30:00 230.3 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

06/15/2022 01:09 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_