
Midmark M11 - v1.0.5
Total Cycles: 645

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 170F
CHAMBER PRESS 3.0PSI
** ITEMS NOT STERILE **

12/16/2021 01:41 PM