
Midmark M11 - v1.0.5
Total Cycles: 2343

Name: RITTER2
Sterilizer ID: V2313228
05/05/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.0 F 0.0
2:00 82.2 F 0.0
4:00 121.9 F 0.0
6:00 124.8 F 0.0
8:00 171.5 F 2.8
10:00 197.0 F 6.9
12:00 214.7 F 11.4
14:00 223.8 F 5.1
16:00 230.7 F 9.2
18:00 240.3 F 11.6
20:00 240.0 F 11.0
22:00 237.3 F 9.1
24:00 251.4 F 15.9
26:00 263.0 F 22.7
28:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.7 F 28.6
2:00 271.8 F 28.8
3:00 271.9 F 28.8
4:00 271.8 F 28.7

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 232.6 F 0.0
Dry Cycle Aborted

FILLING: 1:06
HEATING: 28:03
STERILIZING: 4:00
VENTING: 2:44
DRYING: 0:12
TOTAL CYCLE: 00:36:05

05/05/2023 09:32 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____