
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001947
11:03 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:28
136.4 C 225kPa 4:28
136.3 C 225kPa 4:58
136.2 C 225kPa 5:28
136.3 C 223kPa 5:58
136.1 C 223kPa 6:28
136.1 C 223kPa 6:58
136.2 C 224kPa 7:28
136.2 C 225kPa 7:58
136.2 C 225kPa 7:58
VENTING 7:58
AIR DRYING 8:51
CYCLE COMPLETE 12:26

Digital Signature #
8122817018637343
