
Midmark M11 - v1.0.0
Total Cycles: 3486

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 190.3 F 2.9
4:00 192.1 F 7.0
6:00 220.5 F 12.9
8:00 226.2 F 7.9
10:00 239.2 F 13.5
12:00 239.8 F 11.0
14:00 237.7 F 9.4
16:00 252.0 F 16.3
18:00 263.7 F 23.1
20:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.5 F 28.3
3:00 272.6 F 29.0
4:00 271.8 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 223.2 F 0.0
10:00 233.3 F 0.0
15:00 225.1 F 0.0
20:00 223.0 F 0.0
25:00 224.1 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

07/05/2023 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____