

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2690

Name: RITTER2  
Sterilizer ID: V2313228  
07/24/2023 03:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 157.7 F 1.0  
4:00 173.0 F 4.6  
6:00 193.2 F 8.4  
8:00 209.1 F 12.5  
10:00 220.1 F 6.2  
12:00 230.4 F 10.5  
14:00 232.8 F 8.4  
16:00 242.2 F 12.7  
18:00 241.5 F 11.5  
20:00 253.2 F 17.4  
22:00 263.7 F 23.6  
24:00 272.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.6 F 29.7  
2:00 271.7 F 28.9  
3:00 272.0 F 29.6  
4:00 272.5 F 29.4

Min 271.3 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 211.1 F 0.0  
10:00 210.0 F 0.0  
15:00 208.0 F 0.0  
20:00 210.0 F 0.0  
25:00 212.7 F 0.0  
30:00 212.5 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:03:03  
CYCLE COMPLETE

07/24/2023 04:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_