
Midmark M11 - v1.0.5
Total Cycles: 3062

Name: RITTER2
Sterilizer ID: V2313228
10/30/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.7 F 0.0
2:00 123.3 F 0.3
4:00 136.1 F 2.1
6:00 167.1 F 5.4
8:00 192.0 F 9.3
10:00 231.5 F 13.4
12:00 218.4 F 7.0
14:00 232.5 F 11.6
16:00 232.3 F 8.1
18:00 243.1 F 12.5
20:00 243.9 F 12.6
22:00 255.9 F 18.9
24:00 266.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 28.8
3:00 272.4 F 29.6
4:00 271.6 F 28.9

Min 271.3 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 202.2 F 0.0
10:00 186.2 F 0.0
15:00 186.0 F 0.0
20:00 182.8 F 0.0
25:00 178.6 F 0.0
30:00 175.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

10/30/2023 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____