
Midmark M11 - v1.0.5
Total Cycles: 3350

Name: RITTER2
Sterilizer ID: V2313228
02/02/2024 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.3 F 0.0
2:00 124.2 F 0.3
4:00 140.4 F 2.2
6:00 171.1 F 5.4
8:00 187.0 F 9.4
10:00 208.6 F 9.7
12:00 228.7 F 7.7
14:00 238.9 F 12.4
16:00 235.6 F 9.4
18:00 238.8 F 10.0
20:00 247.4 F 14.1
22:00 258.8 F 20.6
24:00 270.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.5
2:00 272.3 F 29.9
3:00 271.6 F 29.2
4:00 272.1 F 29.8

Min 271.3 F 29.1
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 244.1 F 0.0
10:00 193.7 F 0.0
15:00 185.1 F 0.0
20:00 181.9 F 0.0
25:00 179.7 F 0.0
30:00 178.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:04
CYCLE COMPLETE

02/02/2024 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____