

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5032

Name: RITTER 1  
Sterilizer ID: V2178198  
10/29/2024 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.6 F 0.0  
2:00 171.6 F 0.4  
4:00 179.8 F 2.5  
6:00 215.1 F 6.9  
8:00 220.5 F 11.7  
10:00 227.8 F 6.2  
12:00 238.5 F 11.3  
14:00 235.3 F 8.2  
16:00 242.6 F 10.9  
18:00 246.5 F 13.1  
20:00 258.7 F 19.7  
22:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 27.9  
1:00 272.5 F 28.6  
2:00 271.4 F 27.8  
3:00 272.6 F 28.8  
4:00 271.7 F 28.2

Min 271.2 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
5:00 194.3 F 0.0  
10:00 192.2 F 0.0  
15:00 194.3 F 0.0  
20:00 215.3 F 0.0  
25:00 245.1 F 0.0  
30:00 239.3 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 23:00  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:01:30  
CYCLE COMPLETE

10/29/2024 10:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_