
Midmark M11 - v1.0.5
Total Cycles: 1397

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 189.6 F 1.4
4:00 198.9 F 6.8
6:00 216.7 F 12.2
8:00 228.9 F 8.0
10:00 239.7 F 13.5
12:00 241.2 F 11.6
14:00 240.2 F 10.6
16:00 253.4 F 17.3
18:00 264.5 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.3
2:00 271.6 F 29.0
3:00 272.6 F 29.3
4:00 271.4 F 28.8

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 238.1 F 0.0
10:00 225.2 F 0.0
15:00 219.9 F 0.0
20:00 204.4 F 0.0
25:00 200.4 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

08/16/2022 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____