
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007666
14:00 11/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:48
STERILIZING 3:51
136.0 C 220kPa 3:51
136.3 C 224kPa 4:21
136.3 C 224kPa 4:51
136.4 C 225kPa 5:21
136.3 C 224kPa 5:51
136.4 C 222kPa 6:21
136.1 C 220kPa 6:51
136.4 C 225kPa 7:21
136.4 C 225kPa 7:21
VENTING 7:21
AIR DRYING 8:21
CYCLE COMPLETE 38:21

Digital Signature #
E0B9C9DCF2396FA4
