
Midmark M11 - v1.0.0
Total Cycles: 1209

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2021 06:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 81.4 F 0.0
4:00 113.8 F 0.0
6:00 112.3 F 0.1
8:00 143.8 F 0.1
10:00 172.3 F 0.2
12:00 197.7 F 2.2
14:00 212.1 F 5.4
16:00 220.7 F 8.7
18:00 230.6 F 12.3
20:00 231.1 F 7.8
22:00 242.7 F 12.9
24:00 240.1 F 10.6
26:00 236.5 F 8.8
28:00 251.1 F 15.7
30:00 262.6 F 22.2
32:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.5 F 28.9
2:00 271.8 F 28.0
3:00 272.7 F 28.9
4:00 271.8 F 28.0

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 194.6 F 0.0
10:00 202.3 F 0.0
15:00 202.0 F 0.0
20:00 188.0 F 0.0
25:00 189.9 F 0.0
30:00 192.3 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

09/28/2021 07:42 AM
ACCEPTED: X
REJECTED:
Operator: U01