
Midmark M11 - v1.0.0
Total Cycles: 5110

Name: RITTER 1
Sterilizer ID: V2178198
11/19/2024 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.4 F 0.0
2:00 135.4 F 0.0
4:00 156.2 F 0.3
6:00 186.8 F 0.8
8:00 208.6 F 4.5
10:00 220.0 F 8.6
12:00 230.7 F 13.0
14:00 233.1 F 8.9
16:00 240.8 F 10.7
18:00 243.6 F 12.4
20:00 241.6 F 11.1
22:00 254.8 F 17.8
24:00 266.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.1 F 28.8
2:00 272.3 F 28.6
3:00 271.5 F 28.2
4:00 272.5 F 29.0

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 209.0 F 0.0
10:00 219.4 F 0.0
15:00 223.5 F 0.0
20:00 217.1 F 0.0
25:00 224.3 F 0.0
30:00 232.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 25:29
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

11/19/2024 03:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____