

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3113

Name: RITTER2  
Sterilizer ID: V2313228  
11/10/2023 03:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 96.4 F 0.0  
2:00 97.9 F 0.2  
4:00 118.1 F 1.6  
6:00 139.7 F 4.3  
8:00 170.5 F 7.8  
10:00 188.2 F 11.9  
12:00 210.6 F 4.9  
14:00 220.3 F 9.1  
16:00 233.1 F 12.0  
18:00 235.3 F 10.3  
20:00 235.8 F 8.9  
22:00 246.9 F 14.3  
24:00 258.1 F 20.5  
26:00 268.8 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.2  
1:00 272.5 F 29.7  
2:00 271.6 F 29.6  
3:00 271.9 F 29.3  
4:00 272.5 F 29.9

Min 271.3 F 29.1  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 202.7 F 0.0  
10:00 200.7 F 0.0  
15:00 194.7 F 0.0  
20:00 193.2 F 0.0  
25:00 190.5 F 0.0  
30:00 189.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:02  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:05:30  
CYCLE COMPLETE

11/10/2023 04:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_