
Midmark M11 - v1.0.5
Total Cycles: 3239

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 168.0 F 2.2
4:00 185.8 F 6.5
6:00 207.1 F 11.3
8:00 221.4 F 5.9
10:00 230.7 F 10.9
12:00 232.7 F 7.9
14:00 242.4 F 11.8
16:00 245.3 F 13.1
18:00 257.5 F 19.6
20:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.0 F 29.1
2:00 272.9 F 29.4
3:00 272.3 F 29.7
4:00 271.3 F 29.2

Min 271.2 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 208.7 F 0.0
10:00 198.1 F 0.0
15:00 197.9 F 0.0
20:00 195.4 F 0.0
25:00 192.2 F 0.0
30:00 190.6 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

12/20/2023 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____