
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001660
12:48 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:20
STERILIZING 3:27
136.4 C 224kPa 3:27
136.2 C 225kPa 3:57
136.4 C 221kPa 4:27
136.4 C 225kPa 4:57
136.2 C 222kPa 5:27
136.1 C 223kPa 5:57
136.3 C 225kPa 6:27
136.4 C 226kPa 6:57
136.4 C 226kPa 6:57
VENTING 6:57
AIR DRYING 7:45
CYCLE COMPLETE 8:07

Digital Signature #
363C2D8DEE3E6E58
