

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2377

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2022 10:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.6 F 0.0  
2:00 190.6 F 4.1  
4:00 189.0 F 8.9  
6:00 219.5 F 7.3  
8:00 232.5 F 9.4  
10:00 235.0 F 9.0  
12:00 242.9 F 12.7  
14:00 241.8 F 11.5  
16:00 255.9 F 18.6  
18:00 267.9 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 271.9 F 29.0  
2:00 272.7 F 29.0  
3:00 271.5 F 28.4  
4:00 272.0 F 29.0  
  
Min 271.2 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 230.0 F 0.0  
10:00 224.6 F 0.0  
15:00 231.0 F 0.0  
20:00 231.1 F 0.0  
25:00 230.1 F 0.0  
30:00 232.3 F 0.0  
  
Drying Complete

FILLING: 1:59  
HEATING: 19:08  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

09/27/2022 11:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_