
Midmark M11 - v1.0.0
Total Cycles: 1850

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2022 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.0 F 0.0
2:00 199.8 F 5.3
4:00 211.5 F 11.0
6:00 220.8 F 5.8
8:00 235.3 F 11.4
10:00 235.6 F 8.9
12:00 239.1 F 9.5
14:00 249.6 F 14.9
16:00 262.4 F 22.1
18:00 272.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.1
3:00 272.5 F 28.8
4:00 271.8 F 28.1

Min 271.4 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 229.9 F 0.0
10:00 236.7 F 0.0
15:00 231.9 F 0.0
20:00 223.7 F 0.0
25:00 226.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

04/20/2022 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____