
Midmark M11 - v1.0.5
Total Cycles: 3570

Name: RITTER2
Sterilizer ID: V2313228
04/30/2024 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.0 F 0.0
4:00 74.7 F 0.2
6:00 95.5 F 1.6
8:00 133.2 F 3.9
10:00 160.3 F 6.7
12:00 181.8 F 9.8
14:00 201.3 F 13.2
16:00 210.4 F 4.9
18:00 223.9 F 8.3
20:00 230.2 F 12.1
22:00 230.1 F 7.8
24:00 241.5 F 12.6
26:00 238.9 F 10.4
28:00 250.6 F 16.1
30:00 260.6 F 21.8
32:00 270.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 271.5 F 29.3
2:00 272.4 F 29.6
3:00 271.5 F 28.9
4:00 271.8 F 29.6

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 223.5 F 0.0
10:00 185.7 F 0.0
15:00 180.3 F 0.0
20:00 175.7 F 0.0
25:00 173.8 F 0.0
30:00 183.8 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:50
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:11:02
CYCLE COMPLETE

04/30/2024 10:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____